

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Addiese: COMMISSIONER FOR PATENTS P O Box 1430 Alexandra, Virginia 22313-1450 www.wepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO. CONFIRMATION NO.		
10/508,445	09/16/2004	Joseph P Orban III	2786	6710	
Covidien	7590 01/09/20	7590 01/09/2009		EXAMINER	
60 Middletov		SMITH, FANGEMONIQUE A			
North Haven,	, CT 06473		ART UNIT	PAPER NUMBER	
			3736		
			MAIL DATE	DELIVERY MODE	
			01/09/2009	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Application No.	Applicant(s)	Applicant(s)		
10/508,445	ORBAN, JOSEPH P			
Examiner	Art Unit			
Fangemonique Smith	3736			

arned patent term adjustment. See 37 CFR 1.704(b).
--

Office Action Summary	Examiner	Art Unit					
	Fangemonique Smith	3736					
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply							
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING D. - Extensions of time may be available under the provisions of 37 CFR 1.1 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is generally entered period for reply with the sate or extended period for reply with the sate or extended period for reply with the sate of the sate of the period for reply with the sate of t	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tin will apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	N. nely filed the mailing date of this o D (35 U.S.C. § 133).					
Status							
1)⊠ Responsive to communication(s) filed on 10 Oc 2a)□ This action is FINAL. 2b)⊠ This 3)□ Since this application is in condition for allowar closed in accordance with the practice under E	action is non-final. nce except for formal matters, pro		e merits is				
Disposition of Claims							
Al	vn from consideration.						
Application Papers							
9) The specification is objected to by the Examine: 10) The drawing(s) filed on is/are: a) acce Applicant may not request that any objection to the to Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the Ex	epted or b) objected to by the lidrawing(s) be held in abeyance. Section is required if the drawing(s) is obj	a 37 CFR 1.85(a). jected to. See 37 C					
Priority under 35 U.S.C. § 119							
12) ☐ Acknowledgment is made of a claim for foreign a) ☐ All b) ☐ Some * c) ☐ None of: 1. ☐ Certified copies of the priority documents 2. ☐ Certified copies of the priority documents 3. ☐ Copies of the certified copies of the prior application from the International Bureau * See the attached detailed Office action for a list of	s have been received. s have been received in Applicati ity documents have been receive I (PCT Rule 17.2(a)).	on No ed in this National	Stage				
Attachment(s)	(
Notice of References Cited (PTO-892) Notice of Draftsperson's Patent Drawing Review (PTO-948) Information Disclosure Statement(s) (PTO/SS/CE) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal P 6) Other:	ate					

Office Action Summary

Part of Paper No./Mail Date 20081207

Application/Control Number: 10/508,445 Page 2

Art Unit: 3736

DETAILED ACTION

 This Office Action is responsive to the Request for Continued Examination filed on October 10, 2008. Examiner acknowledges the amendment of claims 1, 19, 28, 31 and 33.
 Claims 1-35 are pending.

Claim Rejections - 35 USC § 103

- The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all
 obviousness rejections set forth in this Office action:
 - (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- Claims 1-5, 7-9, 12-16 and 18-35 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cochran et al. (U.S. Patent Number 5,215,521) in view of Bardes et al. (U.S. Patent Number 5,149,159).

In regard to claims 1-5, 7-9, 12-16 and 18-35, Cochran discloses a tissue removal device and method comprising an elongate shaft (26) having a distal end and a proximal end, wherein the elongate shaft defines a longitudinal axis. Cochran et al. further disclose a bag support (26) defining an opening, the bag support being collapsible and expandable to open and close the opening. The device includes a conical shaped bag (22) operatively connected to the bag support. The bag (22) has an open first end operatively secured to the bag support and a closed end. The bag support in a collapsed position, substantially closes the first end of the bag. Cochran et al. disclose a tube (16) for receiving the shaft, bag support and bag. The Cochran et al. device has folds approximately parallel with the longitudinal axis as shown in Figure 3. The

Page 3


Art Unit: 3736

bag assembly includes a sheath (20) disposed about the bag assembly and the sheath is formed of a plastic lattice construction, which is inherently capable of having a weakness for breaking away from the bag (col. 9, lines 40-50). Cochran et al. additionally disclose an actuation system operatively connected to the sheath in order to facilitate removal of the sheath from the bag assembly (col. 10, lies 28-35). The actuation system includes a tear line formed in the sheath to facilitate tearing of the sheath (col. 10, lines 1-5 and 28-35). Also included with the actuation member is an expandable member (24, 26) positioned within the bag assembly. The expandable member is initiated remotely from the bag. Cochran et al. disclose a tissue removal device wherein the bag (22) in conical in shape and has a proximal edge and a distal edge. The bag includes a slot formed in the vicinity of the bag support near the proximal edge to enable the diameter of the bag to be adjusted. The device disclosed by Cochran et al. further includes a control line (26). Upon use, the Cochran et al. device includes method steps of folding the bag such that the proximal edge is capable of crossing the distal edge (Fig.3) and the folds are approximately parallel to the longitudinal axis. Use of the Cochran et al. device further includes method steps of closing the upper end of the bag using the support, transforming the bag from an open to a closed position. The bag is folded onto itself and the folded bag is placed into a sheath. Cochran et al. disclose the features of the Applicant's invention as described above. Although the Cochran et al. device is capable of folding transversely, Cochran et al. do not specifically disclose this feature. Bardes et al. disclose a sample retrieval instrument having a bag having at least one transverse fold. The device disclosed by Bardes et al. includes a pouch having a shape for the collection and removal of samples. The pouch includes a first and second edge with bag

Application/Control Number: 10/508,445

Art Unit: 3736

folds including at least one transverse fold extending from the first edge toward the second edge such that an angle less than 90 degrees is defined relative to the longitudinal axis (Figure 7).



The fold of the bag increases durability of the bag as taught by Bardes et al. It would have been obvious to one having ordinary skill in the art at the time the Applicants' invention was made to modify a tissue removal device and method, similar to that disclosed by Cochran et al., to include a bag which folds transversly, similar to that disclosed by Bardes et al., to find another way to collapse the bag while maintaining the sample within the bag.

4. Claims 1, 4 and 6 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cochran et al. (U.S. Patent Number 5,215,521) in view of Bardes et al. (U.S. Patent Number 5,149,159) and in further view of Summer (U.S. Patent Number 5,899,694).
In regard to claims 1, 4 and 6, the combined references of Cochran et al. and Bardes et al. disclose a tissue removal device and method, the combination fails to specifically disclose the sheath being fabricated from a flexible heat-shrinking polymer. Summer discloses a sheath

Art Unit: 3736

5.

fabricated from a flexible heat-shrinking polymer. It would have been obvious to one having ordinary skill in the art at the time the Applicants' invention was made to modify a tissue removal device and method, similar to that disclosed by the combined references of Cochran et al. and Bardes et al., to include a sheath made from a heat-shrinking polymer, similar to that disclosed by Summer, to allow the sleeve to assume other shapes including a bent shape if desired.

Claims 10 and 11 are rejected under 35 U.S.C. 103(a) as being unpatentable over


- Cochran et al. (U.S. Patent Number 5,215,521) in view of Bardes et al. (U.S. Patent Number 5,149,159) and in view of Snow et al. (U.S. Patent Number 6,402,722).

 In regard to claims 10 and 11, the combined references of Cochran et al. and Bardes et al. disclose the features of the Applicant's invention as described above. The combined references do not disclose including a cord operatively connected to the tear line for facilitating the tearing of the sheath. Snow et al. disclose a cord (18) operatively connected to a tear line for facilitating the tearing of a sheath along the tear line. It would have been obvious to one having ordinary skill in the art at the time the Applicants' invention was made to modify a tissue removal device and method, similar to that disclosed by the combined references of Cochran et al. and Bardes et al., to include a cord operatively connected to the tear line, similar to that disclosed by Snow et al., to have better control over when the sheath is separated from the rest of the device.
- Claim 17 is rejected under 35 U.S.C. 103(a) as being unpatentable over Cochran et al.
 (U.S. Patent Number 5,215,521) in view of Bardes et al. (U.S. Patent Number 5,149,159) and in further view of Kammerer et al. (U.S. Patent Number 5,480,404).

Application/Control Number: 10/508,445

Art Unit: 3736

In regard to claim 17, the combined references of Cochran et al. and Bardes et al. discloses a tissue removal device and method. The combination discloses a tissue removal device wherein the bag is conical in shape and has a proximal edge and a distal edge. The bag includes a slot formed in the vicinity of the bag support near the proximal edge to enable the diameter of the bag to be adjusted. Although the combined references disclose the features of the Applicant's invention as described above, the combination does not disclose having a bag with a trapezoidal shape. Kammerer et al. disclose a surgical tissue retrieval instrument having a bag having at least one transverse fold. The device disclosed by Kammerer et al. includes a pouch having a trapezoidal shape for the collection and removal of tissue samples. The pouch includes a first and second edge with bag folds including at least one transverse fold extending from the first edge toward the second edge such that an angle less than 90 degrees is defined relative to the longitudinal axis (Figures 3 and 4).



It would have been obvious to one having ordinary skill in the art at the time the Applicants' invention was made to modify a tissue sampling device, similar to that disclosed by the combined references of Cochran et al. and Bardes et al., to include a bag with a trapezoidal shape, similar to that disclosed by Kammerer et al., to provide a larger collection area for the sample being collected.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Fangemonique Smith whose telephone number is (571)272-8160. The examiner can normally be reached on Mon - Fri 8am - 4:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on 571-272-4726. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/508,445

Page 8

Art Unit: 3736

Supervisory Patent Examiner, Art Unit 3736